(b) Classification. Class I. The device is exempt from the premarket notification procedures in part 807, subpart E of this chapter. The device also is exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of \$820.180, with respect to general requirements concerning records, and \$820.198, with respect to complaint files.

[52 FR 33355, Sept. 2, 1987, as amended at 53 FR 35607, Sept. 14, 1988]

#### §886.5915 Optical vision aid.

- (a) *Identification*. An optical vision aid is a device that consists of a magnifying lens with an accompanying AC-powered or battery-powered light source intended for use by a patient who has impaired vision to increase the apparent size of object detail.
- (b) Classification. Class I. The AC-powered device and the battery-powered device are exempt from the premarket notification procedures in subpart E of part 807 of this chapter. The battery-powered device is also exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of \$820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

 $[55\ FR\ 48443,\ Nov.\ 20,\ 1990,\ as\ amended\ at\ 59\ FR\ 63014,\ Dec.\ 7,\ 1994]$ 

# §886.5916 Rigid gas permeable contact lens.

- (a) Identification. A rigid gas permeable contact lens is a device intended to be worn directly against the cornea of the eye to correct vision conditions. The device is made of various materials, such as cellulose acetate butyrate, polyacrylate-silicone, or silicone elastomers, whose main polymer molecules generally do not absorb or attract water.
- (b) *Classification.* (1) Class II if the device is intended for daily wear only.
- (2) Class III if the device is intended for extended wear.
- (c) Date PMA or notice of completion of a PDP is required. As of May 28, 1976, an approval under section 515 of the act is required before a device described in

paragraph (b)(2) of this section may be commercially distributed. See §886.3.

[52 FR 33355, Sept. 2, 1987, as amended at 59 FR 10284, Mar. 4, 1994]

# §886.5918 Rigid gas permeable contact lens care products.

- (a) *Identification*. A rigid gas permeable contact lens care product is a device intended for use in the cleaning, conditioning, rinsing, lubricating/rewetting, or storing of a rigid gas permeable contact lens. This includes all solutions and tablets used together with rigid gas permeable contact lenses.
- (b) Classification. Class II (Special Controls) Guidance Document: "Guidance for Industry Premarket Notification (510(k)) Guidance Document for Contact Lens Care Products."

[62 FR 30987, June 6, 1997]

#### §886.5925 Soft (hydrophilic) contact lens.

- (a) *Identification.* A soft (hydrophilic) contact lens is a device intended to be worn directly against the cornea and adjacent limbal and scleral areas of the eye to correct vision conditions or act as a therapeutic bandage. The device is made of various polymer materials the main polymer molecules of which absorb or attract a certain volume (percentage) of water.
- (b) *Classification*. (1) Class II if the device is intended for daily wear only.
- (2) Class III if the device is intended for extended wear.
- (c) Date PMA or notice of completion of a PDP is required. As of May 28, 1976, an approval under section 515 of the act is required before a device described in paragraph (b)(2) of this section may be commercially distributed. See § 886.3.

 $[52\ FR\ 33355,\ Sept.\ 2,\ 1987,\ as\ amended\ at\ 59\ FR\ 10284,\ Mar.\ 4,\ 1994]$ 

# §886.5928 Soft (hydrophilic) contact lens care products.

(a) *Identification*. A soft (hydrophilic) contact lens care product is a device intended for use in the cleaning, rinsing, disinfecting, lubricating/rewetting, or storing of a soft (hydrophilic) contact lens. This includes all solutions